

EU Declaration of Conformity (DoC)

We, iHealth Labs, Inc.
120 San Lucar Ct., Sunnyvale, CA 94086, USA

European Representative: iHealthLabs Europe
36 Rue de Ponthieu, 75008, Paris, France

declare under our sole responsibility that the product:

product name: Wireless Blood Pressure Monitor
trade name: iHealth
type or model: BP5
relevant supplementary information: NA

to which this declaration is related, is, when used as specified, in conformity with the technical requirements of the standards and the provisions of the essential requirements of the Directive detailed below.

Directive: Radio equipment Directive, RED 2014 / 53 / EU

The product is in conformity with the following standards and/or other normative documents:

HEALTH and SAFETY(Art. 3(1)(a)):

EN 62479: 2010 Assessment of the compliance of low power electronic and electrical equipment with the basic restrictions related to human exposure to electromagnetic fields (10 MHz to 300 GHz)

EMC (Art. 3(1)(b)):

EN 301489-1 V2.1.1 (2017) Electromagnetic compatibility (EMC) standard for radio equipment and services; Part 1: Common technical requirements; Harmonised Standard covering the essential requirements of articles 3.1 (b) of Directive 2014/53/EU and the essential requirements of article 6 of Directive 2014/30/EU

EN 301489-17 V3.1.1(2017) Electromagnetic Compatibility (EMC) standard for radio equipment and services; Part 17: Specific conditions for Broadband Data Transmission Systems; Harmonised Standard covering the essential requirements of articles 3.1 (b) of Directive 2014/53/EU

SPECTRUM (Art. 3(2)):

EN 300328 V2.1.1 (2016) Wideband transmission systems; Data transmission equipment operating in the 2.4 GHz ISM band and using wide band modulation techniques; Harmonised Standard covering the essential requirements of articles 3.2 of Directive 2014/53/EU

iHealth

Connected to my Health

Place and date of issue (of this DoC): Ming Cong Tianjin 13th Jun. 2018
On behalf of Company name : iHealth Labs, Inc.
Name (in print): Ming Cong
Title: Director of Device